
The notified body

**IBEXU- INSTITUT FÜR SICHERHEITSTECHNIK GMBH INSTITUT AN DER TECHNISCHEN UNIVERSITÄT -
BERGAKADEMIE FREIBERG NB 0637**

has carried out the type examination and issued the following certificate **IBExU09ATEX1164**.

The notified body

TÜV NORD CERT GmbH NB 0044 is responsible for QA-monitoring.

The product has been subjected to the conformity assessment procedures "Internal production control"
(Module A)

and in accordance with the ATEX Directive

- a "type examination" (module B) and
- "Conformity to type based on quality assurance of the production process" (Module D).

The manufacturer is responsible for issuing this declaration of conformity with regard to the fulfilment of the essential requirements and the preparation of the technical documentation.

Manufacturer

FISCHER Mess- und Regeltechnik GmbH


Bielefelder Str. 37a
32107 Bad Salzufen, Germany


Tel. +49 (0)5222 974 0

**The devices bear the
following marking:**

CE 0044

IBExU09ATEX1164

 II 1/2G Ex ia IIC T4 Ga/Gb

 II 2D Ex ia IIC T80°C Db

T_{amb} : -10°C bis 60°C

**Bad Salzufen
07 Jan 2025**



T. Malischewski
Managing Director

